

UKCA (UK Conformity Assessed) Declaration of Conformity

(DOC-01225 Rev. 0)

Manufacturer's Name: Excelitas Canada Inc.
Manufacturer's Address: 2260 Argentia Road, Mississauga, Ontario L5N 6H7 Canada

Authorized UK Representative: Qioptiq Photonics Ltd.
 Mitchell Point, Ensign Way, Hamble, Hampshire
 SO314RF, United Kingdom
 Tel: (+44) 2380-744500

Declares that the product: **Equipment Type:** **Laser-LED Microscopy Illumination Device**
Trade Name/Model No.: **X-Cite XYLIS II / XT730L, XT730S**

Conforms to the following regulations:	Safety	0009/21 of 1 January 2021, (S.I. 2016/1101)
	EMC	0007/21 of 1 January 2021, (S.I. 2016/1091)
	RoHS	0037/21 of 1 January 2021, (S.I. 2012/3032)

Conforms to the following Standards:

Product Safety

CSA C22.2 No. 61010-1-12:2012/A1:2018-11 UL 61010-1:2012/R:2018-11 IEC 61010-1:2010/AMD1:2016 EN 61010-1:2010/A1: 2019	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements
IEC 62471:2006 (First Edition) EN 62471:2008	Photobiological Safety of Lamps and Lamp Systems
IEC 60825-1:2014 (Third Edition) EN 60825-1: 2014+A1:2021	Safety of laser products - Part 1: Equipment classification and requirements
21 CFR 1040.10 and 1040.11	FDA/CDRH Code of Regulations, Title 21 – Food and Drugs, Section 1040.10 – Laser Products and 1040.11 – Specific Purpose Laser Products, except for deviations pursuant to Laser Notice No. 50, dated June 24, 2007, or No. 56, dated May 8, 2019

Electromagnetic Compatibility

IEC 61326-1:2012 / EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements – Part 1 General requirements
CISPR 11:2009 +A1:2010 / EN 55011:2009 +A1:2010, Class A, Group 1	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement
FCC Part 15, Subpart B – Class A:2020	Unintentional Radiators
ICES-003, Issue 6, Class A:2020	Information Technology Equipment (Including Digital Apparatus) – Limits and Methods of Measurement

UKCA (UK Conformity Assessed) Declaration of Conformity

(DOC-01225 Rev. 0)

I, the undersigned, hereby declare that the equipment specified above conforms to the above regulations and standards specified.

Daoud Attayi

2023-02-07

Signature

Date

Full Name: Daoud S. Attayi
Position: Regulatory Specialist