

## UKCA (UK Conformity Assessed) Declaration of Conformity

(DOC-01036 Rev.0)

**Manufacturer's Name:** Excelitas Canada Inc.  
**Manufacturer's Address:** 2260 Argentia Road, Mississauga, Ontario L5N 6H7 Canada

**Authorized UK Representative:** Qioptiq Photonics Ltd.  
Mitchell Point, Ensign Way, Hamble, Hampshire  
SO314RF, United Kingdom  
Tel: (+44) 2380-744500

**Declares that the product:**                      **Equipment Type:**                      **Laboratory Equipment, Ultraviolet/Visible Radiometer**  
**Trade Name/Model No.:**                      **X-Cite / XR2100 & XP750 Radiometer & Detector**

**Conforms to the following regulations:**

<b>Safety</b>	<b>0009/21 of 1 January 2021, (S.I. 2016/1101)</b>
<b>EMC</b>	<b>0007/21 of 1 January 2021, (S.I. 2016/1091)</b>
<b>RoHS</b>	<b>0037/21 of 1 January 2021, (S.I. 2012/3032)</b>

### Conforms to the following Standards:

#### *Product Safety*

CSA C22.2 No. 61010-1:2012/R:2016-04 UL 61010-1:2012/U2:2016-04 / EN 61010-1:2010 IEC 61010-1:2010 (Third Edition)	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, Part 1: General Requirements
--	--

#### *Electromagnetic Compatibility*

EN 61326-1:2013	Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements, Part 1: General Requirements
CISPR 11:2009 +A1:2010 / EN 55011:2009 +A1:2010, Class A, Group 1	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement
FCC Part 15, Subpart B – Class A	Unintentional Radiators
ISED Canada ICES-003, Issue 6, Class A	Information Technology Equipment (Including Digital Apparatus)

I, the undersigned, hereby declare that the equipment specified above conforms to the above regulations and standards specified.

*Daoud Attayi*

2021-07-19

Signature

Date

**Full Name:** Daoud S. Attayi  
**Position:** Regulatory Specialist