



**LUMEN DYNAMICS**

PUTTING YOU IN CONTROL

An Excelitas Technologies® Company

**Declaration of Conformity**

<b>Lumen Dynamics Part Number</b>	<b>012-63000</b>
<b>Description</b>	<b>Assy, HXP Lamp Module X-Cite120</b>

Lumen Dynamics hereby certifies that the parts or products indicated above are in compliance with:

**EU RoHS** Directive 2011/65/EU of the European Parliament and the Council as of June 8, 2011 on the restriction of the use of certain hazardous substances in the electrical and electronic equipment, as last amended by Commission Delegated Directive (EU)2015/863.


**China RoHS** Management Methods for the Restriction of the Use of Hazardous Substances in Electrical and Electronic Products, as of July 1, 2016.


**To the best of our knowledge, at this time the part or product indicated above is in conformance of all applicable local and global regulations regarding Mercury emissions and use.**

**Mercury Content** The product indicated above is classified as a Metal Halide Lamp and contains a minimal amount of Mercury, below 1000 ppm or 0.1% maximum concentration by weight in homogenous material.

**Handling/Safety** Handling, Installation and Safety Precaution information is included in each part/product, please refer to product insert.

**Disposal** Please check with your local authority and dispose in accordance with applicable disposal laws.  
Disposal of this product correctly will help save valuable resources and prevent any negative effects on human health and the environment, which could otherwise arise from inappropriate waste handling.

  
\_\_\_\_\_  
Signature

  
\_\_\_\_\_  
Date

**Full Name:** Rebecca Birtch  
**Position:** Quality Assurance & Quality Systems Manager



www.excelitas.com/x-cite  
x-cite@excelitas.com

www.excelitas.com/omnicure  
omnicure@excelitas.com

Lumen Dynamics Group Inc. 2260 Argentea Road, Mississauga, Ontario, L5N 6H7 CANADA  
Telephone: +1 905 821-2600 • Toll Free (USA and CAN): +1 800 668-8752 • Fax: +1 905 821-2055